

# AGFA X-Ray Film System Certificate No. BAM/ZBA/024/2000 3<sup>rd</sup> Revised Version



Hereby it is certified by the BAM Certification Body that the industrial X-ray film system with the designation

12200 Berlin, Germany T: +49 30 8104-0 F: +49 30 8104-7 2222

# Film Agfa Structurix D 5 processed in the

#### developer Agfa Structurix G135

## automatic film processing with immersion time of 100 seconds, developer temperature 28°C

of the manufacturer

#### Agfa-Gevaert n.v., Mortsel/Belgium

meets the requirements of the film system class **C 4** for industrial x-ray films for radiographic testing according to DIN EN ISO 11699-1:2012 as well as the requirements of the ASTM system class **I** according to ASTM E 1815-08 (2013).

### Certificate holder according to certification contract No. BAM-ZBA-0010-2005-GE:

### GE Industrial Belgium BVBA Division Inspection Technologies

The certification was performed on the basis of the above mentioned contract according to the requirements of the ISO/IEC 17065:2012 standard and comprises a type examination with yearly periodical surveillance of the products and quarterly surveillance of the manufacturing process at Agfa-Gevaert n.v., Mortsel, Belgium (BAM Certification Systems II a and II b). The products certified by BAM may be labelled with the BAM certification mark "BAM Geprüft und überwacht" and/or "BAM Certified and under surveillance" together with the certificate number.

The certificate is based on the test report No. VIII.31/6117 of 2000-09-27 as well as on procedures nos. BZS-GS/179/10 and BZS-GS/059/15.

The certificate is valid until October 9, 2020.

for Bundesanstalt für Materialforschung und -prüfung (BAM)

Unter den Eichen 87,12205 Berlin, 2015-10-10

Dr. R. Schmidt BAM Certification Body

Distribution list: 1st Clien

Dr. U. Ewert

Assessor (Technical expert)

2<sup>nd</sup> BAM Certification Body

The BAM Certification Body has been accredited according to standard ISO/IEC 17065:2012 by the DAkkS (Deutsche Akkreditierungsstelle GmbH). The accreditation is valid for the scope given in certificate D-ZE-11075-21-00.

This certificate may only be published in full wording and without any additions. A revocable written consent shall be obtained from BAM beforehand for any amended reproduction or the publication of any excerpts. The German version is legally binding, except an English version is issued exclusively. Place of jurisdiction is Berlin.